



## Certificate

Herewith we, OMRON HEALTHCARE Co., Ltd. certify JPN500 (HEM-7123-AP3) is substantially equivalent to BP765 (HEM-7311-ZSA)\* with HEM-RML31 cuff which is clinically validated employs the same algorithm as the predicate device for determining blood pressure.

\*Hakuo Takahashi, Masamichi Yoshika and Toyohiko Yokoi. Validation of two automatic devices for the self-measurement of blood pressure according to the ANSI /AA MI/IS O81060-2:2009 guidelines: the Omron BP765 (HEM-7311-ZSA) and the Omron BP760N (HEM-7320-Z). *Vascular Health and Risk Management* 2015:11 49-53.

Very Truly Yours,

March, 11, 2019

A handwritten signature in blue ink, reading "Kazunobu Itonaga", is written over a horizontal line.

Kazunobu Itonaga

General Manager

Product Development Department

Product Development Strategy HQ