



Certificate

Herewith we, OMRON HEALTHCARE Co., Ltd. certify the blood pressure monitor JPN750(HEM-7157-AP3) is substantially equivalent to HEM-9200T (Japanese model)* with HEM-RML31 cuff which is clinically validated employs the same algorithm and the same components as the predicate device for determining blood pressure.

*Hakuo Takahashi, Kanako Saito and Nobuki Yakura. Validation of Omron HEM-9200T, a home blood pressure monitoring device for the upper arm, according to the American National Standards Institute/Association for the Advancement of Medical Instrumentation/International Organization for Standardization 81060-2:2013 protocol. *Journal of Human Hypertension* 09 April 2021

Very Truly Yours,

Jan 9, 2024

A handwritten signature in blue ink, appearing to read 'ash', is positioned above a horizontal line.

Takanori Nishioka
Cardiovascular Product Development Department
General Manager